

BIOCOMPATIBILITY CERTIFICATE

Testmaterial: ndd Spirette 2050-0

Supplier: ndd Medizintechnik AG
Technoparkstrasse 1
8005 Zürich
Switzerland

Studies performed: The following studies were performed in order to determine the biocompatibility of the product according to ISO 10993-1:


**CYTOTOXICITY
INTRACUTANEOUS IRRITATION
LOCAL LYMPH NODE ASSAY**

Results: The product did not show any adverse effects in the performed studies. Therefore, the biocompatibility of the test material has been proven.

BSL BIOSERVICE Scientific Laboratories GmbH Munich

Behringstraße 6/8

D-82152 Planegg



Dr. Daniela Stelter

Biological Safety Testing

Date: 27 April 2009

