This document lists all EasyOne Air releases.

Some versions have not been distributed or were distributed only among selected customers. The changes are cumulative, meaning that changes introduced in previous versions are also included in later versions.

If you are using the EasyOne Connect software together with EasyOne Air, please also make sure that your copy of the EasyOne Connect software is up to date.

File EasyOneAir-VersionHistory-V13R.docx

Version	Description	
Date		
V1.17.0	General improvements:	
November 2021	<ul> <li>Added option in settings to show %predicted values for ratio parameters</li> </ul>	
l l	<ul> <li>Updated electronic instructions for use to V04</li> </ul>	
Compatible with		
EOC PC Software	• Bug fix:	
V3.09.02.04	<ul> <li>Fixed issue with Cal-check: device sometimes freezes during flow baseline</li> </ul>	
and higher	setting	
V1.16.0	Spirometry improvements:	
June 2021	$_{\odot}$ User can switch between 2005 and 2019 spirometry standard for FVC/FVL	
Compatible with	test.	
EOC PC Software	• Updated quality grading for SVC test according to 2019 standard.	
V3.09.00.x and	<ul> <li>Reports now conform to ATS/ERS Spirometry Update 2019 and ATS</li> </ul>	
higher	Recommendation for Standardized PFT Reports 2017.	
	• Parameter FEV1/VC is now reported, provided that an SVC maneuver was	
	performed within 12-hours prior to the forced maneuver.	
	<ul> <li>Calibration check: Error tolerance is reduced from 3.5% to 3%.</li> <li>Order handling: EOA new provides a warning message at the start of a regular.</li> </ul>	
	<ul> <li>Order handling: EOA now provides a warning message at the start of a regular test, if an order is pending for the same patient.</li> </ul>	
	<ul> <li>Tests can now be recalculated to reflect the updated spirometry configurations such as predicted set or interpretation method.</li> </ul>	
	Bug fixes:	
	<ul> <li>Opening report with no test provokes an exception.</li> <li>An unexpected error happens during test recalculation.</li> </ul>	
	<ul> <li>Volume-time predicted range is incorrectly displayed on VT graphs.</li> </ul>	
V1.15.0	<ul> <li>On-device operator's manuals:</li> </ul>	
July 2020	• Updated to version 3.	
. ,	<ul> <li>Added Russian and Polish translation.</li> </ul>	
Compatible with	General improvements:	
EOC PC Software	<ul> <li>Added warning when storage is full.</li> </ul>	
V3.07.00.x and	<ul> <li>Reduced maximum audio output volume.</li> </ul>	
higher	• Bug fix:	
	<ul> <li>Unexpected error after deleting a patient belonging to an open order.</li> </ul>	
V1 14 0		
<b>V1.14.0</b> March 2020	Additional Language Support	
	<ul> <li>Added Polish and Russian software languages.</li> </ul>	
Compatible with	<ul> <li>Added some missing translations for all languages.</li> </ul>	
EOC PC Software	Spirometry improvements	
V3.07.00.x and	<ul> <li>Added new predicted set: Kainu (Finland) 2016</li> </ul>	
higher	Bug Fixes     Cannot add any nationt on 20 February	
	<ul> <li>Cannot add any patient on 29 February.</li> <li>SVC/MVV: respiration frequency reported in wrong units.</li> </ul>	
V1.13.0		
December 2019	<ul> <li>Improved cybersecurity         <ul> <li>New dedicated security settings menu.</li> </ul> </li> </ul>	
Compatible with	<ul> <li>Option in security settings to disable printing to PDF.</li> <li>Option in security settings to enforce strong password rules.</li> </ul>	
EOC PC Software	<ul> <li>Modifying date and time now requires admin privilege.</li> </ul>	
V3.06.00.23	<ul> <li>Password is always hidden, with the option to toggle the visibility.</li> </ul>	
and higher	<ul> <li>General improvements</li> </ul>	
	<ul> <li>Generat improvements</li> <li>Enforce unique user IDs when adding and editing users.</li> </ul>	
	<ul> <li>Show user ID, when configured, also on EasyOne Connect report printouts.</li> </ul>	
	<ul> <li>O Prevent removing the last system administrator.</li> </ul>	
	<ul> <li>Rearrange the items in the settings menu for better usability.</li> </ul>	
	<ul> <li>Show lightning bolt beside the battery icon whenever the device is connected</li> </ul>	
	I Show agrands bou beside the battery for whenever the device is connected	

File EasyOneAir-VersionHistory-V13R.docx

14 40 0	
<b>V1.12.0</b> July 2019	<ul> <li>Bug fixes         <ul> <li>Predicted value calculation for SEPAR 2013 is wrong when no patient weight</li> <li>is extended</li> </ul> </li> </ul>
Compatible with EOC PC Software	<ul> <li>is entered.</li> <li>Patient's height shows 12 inches on printout.</li> <li>General improvements</li> </ul>
V3.04.00.23 and higher	<ul> <li>Display operator's manuals on the device.</li> <li>Add or revise some translated texts.</li> </ul>
<b>V1.11.0</b> May 2019	<ul> <li>Bug fixes         <ul> <li>Device freezes in world map after selecting region.</li> <li>Test variability and interpretation are cleared after a recalculation.</li> </ul> </li> </ul>
Compatible with EOC PC Software V3.04.00.23 and higher	<ul> <li>General improvements         <ul> <li>Option to delete one or several patients and all related test reports.</li> <li>Option to delete a test session from the session result screen.</li> <li>Improved detection of very weak FVC/FVL maneuvers.</li> <li>Device issues a warning if the system date is incorrect.</li> <li>Improved online screen graph for SVC.</li> </ul> </li> </ul>
V1.10.2	Bug fixes
November 2018	<ul> <li>Flow tube detection fails on very few devices.</li> </ul>
Compatible with EOC PC Software V3.04.00.23 and higher	• FVC with patient of ethnicity "other" yields wrong predicted values (already fixed in V1.9.1).
<b>V1.10.1</b> September 2018	<ul> <li>Additional support for printing to HP printers (PCL3GUI printer language)</li> <li>Refer to EasyOne Air printing application note online for more details.</li> </ul>
Compatible with EOC PC Software V3.04.00.23 and higher	<ul> <li>Improved useability         <ul> <li>Improve FVC/FVL coaching message before expiration maneuver.</li> <li>Improve robustness of access to internal storage, software update and flow tube detection.</li> <li>Make software update quicker.</li> </ul> </li> <li>Bug fixes</li> </ul>
	<ul> <li>Bug fixes</li> <li>After a factory reset, set ethnic correction factors and configured predicted to defaults.</li> <li>Fix calculation of FEF75(MEF75) predicted value for females in predicted set AUSTRIA_94.</li> </ul>
V1.9.1	Support Provocation/Challenge Tests in combination with EasyOne Connect
March 2018 Compatible with EOC PC Software V3.04.00.23 and higher	<ul> <li>Perform provocation tests with EasyOne Air.</li> <li>Supports full provocation functionality of EasyOne Connect including customized provocation tests.</li> </ul>
	<ul> <li>Improved connection stability in combination with EasyOne Connect         <ul> <li>Improved stability during online tests.</li> <li>Fixed several issues that could lead to Bluetooth disconnections during online tests.</li> </ul> </li> </ul>
	<ul> <li>Generally improved functionality         <ul> <li>Recalculate the predicted values after changing patient demographics.</li> <li>Show progress bar for long operations.</li> </ul> </li> </ul>
	<ul> <li>Additional Language Support         <ul> <li>Added Swedish software language.</li> </ul> </li> </ul>
	<ul> <li>General Bug fixes         <ul> <li>Solved several minor issues that lead to system crashes.</li> </ul> </li> </ul>

000

Page 3 of 5

File EasyOneAir-VersionHistory-V13R.docx

V1.8.0	Additional Language Support
November 2017	<ul> <li>Added Dutch (NL) software language and Operator's Manual.</li> </ul>
	<ul> <li>Added Brazilian Portuguese (BR) software language and Operator's Manual.</li> </ul>
Compatible with	<ul> <li>General translation improvements.</li> </ul>
EOC PC Software	o General translation improvements.
V3.03.00.11 and higher	- Improved Uceshiltiv and Deporting Canabilities
	Improved Useability and Reporting Capabilities
	<ul> <li>Improved software update process.</li> </ul>
	<ul> <li>Better curve visibility with larger online display of VT curves.</li> </ul>
	<ul> <li>Improved tidal breathing detection.</li> </ul>
	<ul> <li>Improved layout of thermal printer reports.</li> </ul>
	$\circ$ "Lung age" and PEF/PIF in [L/min] are now available on reports.
	<ul> <li>Reorganized database and Bluetooth configurations.</li> </ul>
	- Improved Plustooth Connections
	<ul> <li>Improved Bluetooth Connections         <ul> <li>Improvements in Bluetooth for better connection stability.</li> </ul> </li> </ul>
	o improvements in Bluetooth for better connection stability.
	General Improvements and bug fixes
	<ul> <li>Fixed an issue where in rare cases the device could not be started.</li> </ul>
l l	$\circ$ Fixed an issue where for large database the device stopped working and had to be
	restarted.
	<ul> <li>Fixed an issue where using the "Garcia-Rio 2013 (SEPAR)" predicted values</li> </ul>
	sometimes caused the device to crash.
V4 7 0	
<b>V1.7.0</b> July 2017	Improved printer support and data management:
	<ul> <li>Updates can now be applied via EasyOne Connect (V3.02.00.06).</li> </ul>
Compatible with	
EOC PC Software	
V3.02.00.06	
and higher	<ul> <li>Option to delete the last trial of a spirometry test.</li> </ul>
	Support of new predicted values and ethnic correction:
	• Manually set ethnic correction for Asian, African, Hispanic ethnicities.
	<ul> <li>Added predicted values Dejsomritrutai (Thai2000) for Thailand.</li> </ul>
	$\circ$ Added predicted values Chhabra (2014) for India.
	Additional language support for software:
	$\circ$ Added Italian (IT) software language.
	<ul> <li>Added Danish (DA) software language.</li> </ul>
	Additional Operator's Manual languages:
	<ul> <li>Added Italian (IT) manual</li> </ul>
	<ul> <li>Added Spanish (ES) manual</li> <li>Added Franch (FD) manual</li> </ul>
	<ul> <li>Added French (FR) manual</li> </ul>
	<ul> <li>Added Swedish (SV) manual</li> </ul>
V1.6.2	Improvements
May 10, 2017	<ul> <li>Improved performance of database operations (e.g. to display test list, synchronization</li> </ul>
, , =.	with EasyOne Connect). This fixes a problem when the access gets slower the more data is
, ,	
, ,	stored in the device.
, ,	<ul> <li>stored in the device.</li> <li>Added missing license information</li> </ul>
, , -,	Added missing license information.
, ,	
, ,	Added missing license information.

000

Page 4 of 5

File EasyOneAir-VersionHistory-V13R.docx

V1.6.1 April 28, 2017 (very limited distribution)	<ul> <li>Improvements <ul> <li>More effective storage of curves (compression).</li> <li>OSHA print template.</li> <li>Print status feedback information.</li> <li>Language support improved for Spanish, German &amp; French.</li> <li>Admin authentication added for additional functions (e.g. Factory defaults).</li> <li>Prefered test type that is shown in the main menu is now configurable(preset: FVC).</li> <li>Added link to Open Source licenses.</li> </ul> </li> <li>Bug fixes <ul> <li>Restore functionality did not work.</li> <li>Large databases (&gt; 500 patients) did cause a crash.</li> <li>Device hangs when Australia was selected in the region selection.</li> <li>Device hangs when the date &amp; time has been confirmed very quickly (&lt; 1s) in the first time startup wizard.</li> </ul> </li> </ul>
<b>V1.6.0</b> March 16, 2017 (only selected	EasyOne Connect >= V3.1.0.1  First official distributed release for selected customers
	Compatibility
customers)	EasyOne Connect V3.0.0.3.

## 000